



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/696,932	10/30/2003	James F. McGuckin JR.	1255	1044

7590
NEIL D. GERSHON
REX MEDICAL
1011 HIGH RIDGE RD
Stamford, CT 06905

09/30/2008

EXAMINER

SHAFFER, RICHARD R

ART UNIT	PAPER NUMBER
----------	--------------

3733

MAIL DATE	DELIVERY MODE
-----------	---------------

09/30/2008

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/696,932	Applicant(s) MCGUCKIN ET AL.	
	Examiner Richard Shaffer	Art Unit 3733	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 18 August 2008.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 2-9, 11, 13-18, 21 and 22 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 2-9, 11, 13-18, 21 and 22 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 16 June 2008 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Continued Examination Under 37 CFR 1.114

A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on August 28th, 2008 has been entered.

Drawings

The amended drawing sheet for Figures 4 and 5 filed on June 16th, 2008 is acknowledged and accepted by the examiner. The previous drawing objection is hereby withdrawn.

Claim Rejections - 35 USC § 112

The amendments to the claims filed on June 16th, 2008 are acknowledged and accepted by the examiner. The previous 35 U.S.C. 112, second paragraph rejections are hereby withdrawn.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:


(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 2-7, 9, 11, 13, 14, 16, 18, 21 and 22 are rejected under 35 U.S.C. 103(a) as being unpatentable over Bosma et al (US Patent 6,443,972) in view of Van der Burg et al (US Patent 6,994,092).

Bosma et al disclose a vessel filter (**Figures 5 and 8-9C**) and method (**Figure 1**) comprising: a mounting section (middle section) having a proximal and distal end; two filtering sections (angled portions converging at tubular sections); the filtering sections defined by a plurality of longitudinal struts with ribs connecting adjacent struts; the device is self-expanding Nitinol (**Column 4, Lines 11-15**); the struts are spaced circumferentially about 60 degrees apart; a plurality of vessel engaging members (**27**) and/or roughening (**Column 6, Lines 52-53**); and the struts (in **Figure 8**) have an angled portion with a width less than the width of the straight portion of the mounting section.

Bosma et al disclose all of the claimed limitations except for the specific structure and location of the filtering section with regard to the mounting section. Van der Burg et al teach (**Figure 35; Column 18, Lines 34-55**) that an expandable device for placement against a bodily tubular structure with an S-shape end portion provides an elongated mounting section which improves contact area. It would have been obvious to one having ordinary skill in the art at the time of invention to design the filter of Bosma et al according to the structure taught in Van der Burg et al in order to improve fixation within the blood vessel. In regard to applicant's extensive claim language defining the relationship between the filter section and mounting section, see the Figure below depicting how the structure of Van der Burg et al meets the claimed limitations.

Art Unit: 3733



Claims 8, 15, and 17 are rejected under 35 U.S.C. 103(a) as being unpatentable over Bosma in view of Van der Burg et al and in further view of Gilson et al (US Patent Application Publication 2002/0058911).

The combination of Bosma and Van der Burg et al disclose and teach all of the claimed limitations except for struts out of phase. Gilson et al teach in **Figures 15-22** out of phase struts as an appropriate structure for forming an expandable structure to act as an embolic filter. It would have been obvious to one having ordinary skill in the art at the time the invention was made to design the combination of Bosma and Van der Burg et al with out of phase struts as a matter of mere substitution for providing an expandable embolic filter with predictable results.

Response to Arguments

Applicant's arguments filed June 16th, 2008 have been fully considered but they are not persuasive.

Applicant states that Van der Burg et al lack a distalmost end point of a distal tubular portion proximal of distalmost end points of struts of the mounting section. As shown in the figure in the previous page, at least the two bottom struts on the right are shown distal of the distalmost point of the tubular section as well as the two back struts on the left side being proximal of the proximalmost point of the tubular section.

The examiner would also like to note that in regard to claim 21, due to the claimed invention being directed to a state of the device, the combination of Bosma et al and Van der Burg et al is inherently capable of additional movement of the tubular portions into the center of the device.

Applicant additionally contends that Van der Burg et al lack linear portions of the mounting sections extending axially and radially inward. Again, as previously shown, there are portions that depict a linear strut member.

Applicant also states that Van der Burg et al lack the proximalmost and distalmost end points of the tubular portions axially aligned with a region of the linear portion. As already described in this remarks, there are portions further from the center of the device including a portion of the linear portion therefore axially aligning the distalmost/proximal most tubular end points with the linear portions of the struts.

Conclusion

All claims are drawn to the same invention claimed in the application prior to the entry of the submission under 37 CFR 1.114 and could have been finally rejected on the grounds and art of record in the next Office action if they had been entered in the application prior to entry under 37 CFR 1.114. Accordingly, **THIS ACTION IS MADE FINAL** even though it is a first action after the filing of a request for continued examination and the submission under 37 CFR 1.114. See MPEP § 706.07(b). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Richard Shaffer whose telephone number is (571)272-8683. The examiner can normally be reached on Monday-Friday (7am-5pm).

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on 571-272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3733

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Richard Shaffer/

Examiner, Art Unit 3733

/Eduardo C. Robert/

Supervisory Patent Examiner, Art Unit 3733